

GLOVES

HPV wants to hear from companies with the capacity to supply examination and surgical gloves to support a centrally managed State Supply Chain (SSC). The SSC has been established to ensure essential supplies reach areas with the greatest clinical need. If you have capacity to supply the products outlined below please contact us at suppliers@hpv.zendesk.com. If you would like to offer manufacturing capability to make supplies for the COVID-19 response, please send your offer to medicalPPEsupplies@djpr.vic.gov.au.



If your PPE and equipment are not TGA certified, you will need to begin the application process. Please contact the TGA directly to start this process in parallel with HPV registration. Email: devices@tga.gov.au or call: 1800 141 144.

SCOPE

We are looking for the following types of gloves:

Non-Sterile

- Nitrile
- Non latex
- Suitability for cytotoxic agents.
- Standard length of 240mm and longer at 290mm
- Powder free

SPECIFICATIONS

At a minimum, all products must meet the following standards and testing requirements:

- ISO 2859, ISO 13485
- EN 455: Parts 1 to 4, EN 374: Parts 1 to 3, EN 420, EN 16523-1
- ISO 11193-1:2008: Parts 1 & 2
- ASTM D3578, D6319