

MASKS

HPV wants to hear from companies with the capacity to supply masks to support a centrally managed State Supply Chain (SSC). The SSC has been established to ensure essential supplies reach areas with the greatest clinical need. If you have capacity to supply the products outlined below please contact us at suppliers@hpv.zendesk.com. If you would like to offer manufacturing capability to make supplies for the COVID-19 response, please send your offer to medicalPPEsupplies@djpr.vic.gov.au.



If your PPE and equipment are not TGA certified, you will need to begin the application process. Please contact the TGA directly to start this process in parallel with HPV registration. Email: devices@tga.gov.au or call: 1800 141 144.

SCOPE

We are looking for the following types of single use masks:

N95

Single Use Face Masks, Level 1 ,2 ,3

Mask with visor



SPECIFICATIONS

At a minimum, all products must meet the following standards and testing requirements:

N95

- AS/NZS 1716:2012, AS/NZS 1715:2009
- National Personal Protective Technology Laboratory (NPPTL) tested
- National Institute for Occupational Safety and Health (NIOSH) certified

Single Use Masks

- AS /NZ4381:2015,
- Certified Level 1 - 3 barriers medical face masks are evaluated for resistance to penetration by synthetic blood at minimum velocity specified for bacterial filtration efficiency and differential pressure.
- Level 1 - 80mm Hg , Level 2 - 120mm Hg , Level 3 - 160mm Hg
- ASTM F2101-14 or EN 14683:2014 ASTM F1862 / F1862M-13 or ISO 22609 certified

Mask with Visor

- Level 3 fluid resistant mask with foam band and anti-fog visor 25cm by 10-12cm and ties

Note

Ear loops must not contain latex