

# Small Size N95 Respirator

HPV would like to hear from companies with the capacity to supply small size N95 respirators to support a centrally managed State Supply Chain (SSC). The SSC has been established to ensure essential supplies reach areas with the greatest clinical need.

## Scope

- N95 – P2- PF Respirator
- Internal Nose clip
- Respirator straps – headbands or ear loops

## SPECIFICATIONS

At a minimum, all products must meet the following standards, testing, and packaging requirements:

### Standard requirements

- Therapeutic Goods Administration (TGA) Certified
- AS/NZS 1716:2012
- One or more of the following NIOSH, CE 2797, EN 149:2001 + A1: 2009, FFP2 NR D, AS/NZ 1716: 2012 P2

### Test Reports

- BFE – Bacterial Efficiency Test Report >98%
- Fluid Resistance Test Report in mmHg
- Exhalation Resistance in Pa
- Inhalation Resistance if assembled respirator in Pa

### Packaging & Labelling Requirements

Printed on respirator: IF

- Specific standard
- N95

Printed on packaging:

- Qty/Per Box
- Language English
- ARTG number
- Size
- Fluid resistance e.g. 160 mmHg
- Specific standard

